

Reference your bid dated 26/10/2017 on subject cited above.

It is to inform you that the process of technical evaluation has been completed and this draft Technical Evaluation Report is hereby shared/ announced as a pre-requisite of the procurement process.

The attached proformas indicating marks given to your bid/s is sent to you for your information, clarification and proposing corrections, if any.

This draft report is open to rectification based on typographical/ computational errors and / or omissions & merit based complaint, if any, in order to provide equal opportunity, ensure transparency, fairness and provision of maximum playing field to all bidders for a healthy competition in the public interest.

Please note that during the course of evaluation of your bid/s, certain information was not readily appreciable and thus zero marks were awarded provisionally which is subject to the clearance by T&E committee.

Further to note that this report relates to the items pertaining to category A, B ,E & F and the results of remaining categories i.e. C & D will be announced once their evaluation is completed. Bidders who have applied for leftover categories will be informed accordingly.

You are, therefore, supposed to propose corrections for errors/ omissions in technical evaluation report on our part and not to propose any alterations in the bids already submitted by your firm till 14/12/2017 during office hours.

Evolution Criteria for the Purchase of Category A (Left over Items) (Section-IV of Statement of Requirement) for the financial year 2017-18

Sl#	No of Firms	Firm Name	1	2	Firm Evaluation Parameters									Product Evaluation Parameter								23	24	25				
					3	4	5	6	7	8	9	10	11	12	13	Conformance To Specification			Product International Certification		Product Local Performance				Technical Marks out of 100			
					Personnel/ Human Resource		Workshop facility Testing/Calibration of Equipment			Manufacturer Performance			Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted)			Additional Features of the product		Certificate of US Food and Drug Administration (US FDA)	Certificate of European Community MDD (CE MDD)	Certificate of Japan Industrial Standard (JIS)	Weightage for local Pakistani original manufacturer	One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the public sector medical institution of Pakistan			One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the teaching level private sector medical institution of Pakistan	Warranty Period three year both with spare parts and services & next two year services without parts, from the date of installation	Total Technical Score	Status
					2	2	5	5	5	5	5	5	2	2	2	40	6	No marks, being mandatory parameter			4	3	3	4	100			
1	1	LATIF BROTHERS	Argon Laser model:VITRA multispot make quantel medical france		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Quoted Alternate model
	2	Radiant Medical Pvt Ltd	Argon Laser, Model: Light Las Tnescan, Make: Light Med USA		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
2	1	MEDI LINE TECHNOLOGY	ECT MACHINE MODEL:SPECTRUM 5000M MAKE :MECTA COORPORATION USA		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization is not attested by embassy. Not according to specifications quoted pulse width frequency is 0.5-0.75ms while required is 1.5ms. Stimulus current is fixed to 800mA ,while required is 500 to 800 mA. Charge greater than 5mC on one side
	1	IBS Pharmaceutical	EEG 32 Channel, Model:BWIII EEG System 36 Channel, Make: NeuroVirtual USA		0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization is not from MF Equipmentos (Manufacturer). Date of Embassy attestation of authorization is after closing date for bid i.e 26.10.2017

Evaluation Criteria for the Purchase of Category-B (Left over items) for the Financial Year 2017 -18

Sl No	No of Firms Submitted bid	1	2	3										14										23	24	25				
				Firm Evaluation Parameters										Product Evaluation Parameter													Technical Marks			
				Personnel/ Human Resource		Workshop facility Testing/Calibration of Equipment				Manufacturer Performance				Conformance To Specification		Product International Certification				Product Local Performance								Technical Marks out of 100		
				Simple Technician	Associate Engineer with Diploma of Associate Engineer	Graduate Engineer duly registered with PEC/NASC Electronics with minimum five years' experience	Availability of fully functional works hop in Khyber Pakhtunkhwa	Availability of fully functional works hop at National level	List of related tools for the quoted items on firm's pad	Detail of Spare parts availability on the firm's pad(for the quoted items)	Deduction/certificate of calibration by the firm (for quoted items)	Valid ISO 14001 Environment Certificate	Valid ISO 9001 Quality Management Certificate	Valid ISO 13485 Quality management certificate	Fully compliance with the required specifications as per statement of Requirement (Up to maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted).	Additional features of the product	Certificate of US Food and Drug Administration (US FDA)	Certificate of European Community MDD (CEMDD)	Certificate of Japan Industrial Standard (JIS)	Weightage for local Pakistani original manufacturer	One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the public sector medical institution of Pakistan	One mark for each satisfactory performance certificate for quoted model or at least the previous provided model of equipment from the private sector medical institution of Pakistan	Warranty Period three years both with spare parts and services & next two year services without parts from the date of finalisation				Total Technical Score		Status	Remarks
				2	2	5	5	5	5	5	5	5	5	2	2	2	30	4	4	4	4	4	4	3	3	4	100			
1	1	Paradise Export Co	Air Drill With Cylinder/ Orthopedic Pneumatic Drill with Cylinder Dimend Germany Air driven Handpiece Number 33.326.05	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Not According to specification. Quoted speed is less then the required 900-1250rpm, Cylinder not quoted.	
2	1	JATIF BROTHERS	APPLANATION TONOMETER MODEL :AT-1 MAKE: TAKAGI JAPAN	2	2	5	5	5	5	5	5	0	0	0	30	0	0	0	0	0	0	0	0	0	4	68	Responsive	NIL		
	2	IBS Pharmaceutical	Applanation Tonometer, Model:YZ30R, Make:ARI China.	2	2	5	5	5	5	5	5	0	0	0	30	0	0	0	0	0	0	0	0	0	4	68	Responsive	Nill		
	3	SUDAIS ASSOCIATES	APPLANATION TONOMETER MODEL: AT-R MAKE RAY VISION CHINA	2	2	5	5	5	5	5	5	0	0	0	30	0	0	0	0	0	0	0	0	0	4	68	Responsive	NIL		
3	1	Medco	Autodave Automatic 100 L Model: BKQ-8Z 100(H) Make Biobase China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specification, Pressure is less then Required. Quoted Manual doors	
	2	Clinical Life Inc	Autodave Automatic 100 Liters Model Name: Steri 21 Model number: 40.40.66 Manufacturer: A.J.Costa Origen: Portugal Ro system Model: DM-50-2A	2	2	5	0	5	5	5	5	0	2	2	30	0	0	4	0	0	0	0	0	0	4	71	Responsive	NIL		
	3	Delta Plus Pvt Ltd	Autoclave Automatic 100 Liters Model: Azteca AC-470 Manufacturer: Celttron Medical Technology kft Origen:p Hungry (E.U) RO System: FIZICO 100L Manufacturer: FIZICO-Z.A Engineers Lahore (Dow USA parts)	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.	
	4	RAY TECHNOLOGIES	AUTOCLAVE AUTOMATIC 100 LITERS MODEL:113 HIRAYAMA MAKE:JAPAN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specifications. Separate Steam Generator Quoted	
	5	Total Technology	Autodave Automatic 100 Liters Model: Clinclav 4S Make: Melag Gemany	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specification, Pressure is less then required.	
	6	KASBN	AUTOCLAVE AUTOMATIC 100 LTR MODEL: AS-446 WPA MAKE: SMSM POLAND	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.	

6	1	Vertex Medical (PVT) LTD	Baby Suction Machine Model Basic Make Medela AG Switzerland	2	2	5	5	5	5	5	5	5	5	0	0	2	29	0	4	4	0	0	3	0	4	80	Responsive	Nil	
	2	SUDAIS ASSOCIATES	BABY SUCTION MACHINE MODEL: MS-700 MAKE: M.TECHNOLOGY KOREA	2	2	5	5	5	5	5	5	5	5	0	2	2	30	0	0	4	0	0	0	0	4	76	Responsive	NIL	
	3	KASBN	BABY SUCTION MACHINE MODEL:FV SERIES SM1240 MAKE: ORDISI SPAIN EUROPE	2	2	5	5	5	5	5	5	5	5	0	2	2	30	0	0	4	0	0	2	0	4	78	Responsive	NIL	
	4	IBS Pharmaceutical	Baby Suction Machine, Model:III Negative Pressure, Make: ARI China.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specification Capacity and quantity of Suction Bottle is less then Required.
	5	Al- Kareem Medical Technology	Baby Suction Machine, Model:MEVACS M20, Make:Medist Slovakia	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
7	1	Blazon Diagnostic	BLOOD BAG CENTRIFUGE MODEL: DP-206SR 12 B MAKE: PRESVAC SRL Argentina	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
	2	Medco	Blood Bag Centrifuge Model: BKC-886A Make: Biosse China	2	2	5	5	5	5	5	5	5	0	0	0	30	0	0	0	0	0	0	0	4	68	Responsive	Nil		
	3	IBS Pharmaceutical	Blood Bag Centrifuge, Model: DL-GLM, Make:ARI China.	2	2	5	5	5	5	5	5	5	0	2	2	30	0	0	0	0	0	0	0	4	72	Responsive	Nil		
8	1	Blazon Diagnostic	BLOOD CELL COUNTER MODEL MODEL: CC-1431 CAT 7121431 MAKE: JP SELECTA SPAIN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
9	1	ENDO_KARE	ENT Examination Chair Model SIT 4 OTOPRONT GERMAN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization letter is not Embassy Attested.
	2	MEDIUNE TECHNOLOGY	ENT EXAMINATION CHAIR MODEL:MC 4000A MAKE: MEDSTAR CO LIMITED KOREA	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
	4	Radiant Medical Pvt Ltd	ENT Examination Chair, Model: 45A, Make: Heinemann Germany.	2	2	5	5	5	5	0	5	0	0	2	29	0	0	4	0	0	0	0	0	0	0	64	Responsive	Nil	
10	1	ENDO_KARE	ENT Examination unit with Cautery section Model SMART OTOPRONT GERMAN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization letter is not Embassy Attested.
	3	MEDIUNE TECHNOLOGY	ENT Examination unit with Cautery section MODEL:UE 3000 PLUS MAKE: MEDSTAR CO LIMITED KOREA	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
	2	Radiant Medical Pvt Ltd	ENT Examination unit with Cautery section. Model: Modula 4.5A Europa. Make: Heinemann Germany	2	2	5	5	5	5	0	5	0	0	2	28	0	0	4	0	0	0	0	0	0	0	63	Responsive	Nil	
	3	BioCare	ENT Examination unit with Cautery section. Model: Otlaphari compact plus Dantschke Medizintechnik Germany	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization letter is not Embassy Attested.
11	1	Pakistan Marketing Services	Glucometer Make Viva Chek Laboratories Inc China	2	2	0	5	5	5	5	5	0	0	2	30	0	4	4	0	0	2	2	4	77	Responsive	Nil			
	2	Qazzafi Surgical	Glucometer Model: Easy Gluco Make: Infopia Korea	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
12	1	IBS Pharmaceutical	Gynae Vacuum Suction Apparatus/Vacuum Extractor With Trolley, Model:Z3C.III Mobile Electric Suction Make:ARI China.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specification Capacity of Suction Bottle is less then Required. Suction currents not quoted
	2	Al- Kareem Medical Technology	Gynae Vacuum Suction Apparatus/Vacuum Extractor With Trolley, Model:MEVACS M90, Make: Medist, Slovakia.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
	3	SUDAIS ASSOCIATES	GYNAE VACUUM SUCTION WITH TROLLY MODEL: MS-700 MAKE: M.TECHNOLOGY KOREA	2	2	5	5	5	5	5	5	0	2	2	30	0	0	4	0	0	0	0	0	4	76	Responsive	NIL		
	3	KASBN	GYNAE VACUUM SUCTION MODEL:FV SERIES SM1240 MAKE: ORDISI SPAIN EUROPE	2	2	5	5	5	5	5	5	0	2	2	29	0	0	4	0	0	2	0	4	77	Responsive	NIL			
	4	Vertex Medical (PVT) LTD	Gynae Vacuum Suction Apparatus/ vacuum extractor with accessories Model: DOMINATE FLEX Medela AG Switzerland	2	2	5	5	5	5	5	5	0	0	2	29	0	4	4	0	0	3	0	4	80	Responsive	Nil			
13	1	Al- Kareem Medical Technology	Humidifier Model: AIRnava, Make: WILMed, Germany.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.
14	1	SUDAIS ASSOCIATES	INDIRECT OPHTHALMOSCOPE MODEL:IOP-60 MAKE: RAY VISION CHINA	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specifications, Light intensity is not mentioned in catalogue
	1	LATIF BROTHERS	INDIRECT OPHTHALMOSCOPE MODEL:IOP-ALPIA MAKE: NEITZ JAPAN	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not according to specifications, Light intensity 2000 lux not mentioned
	2	Radiant Medical Pvt Ltd	Indirect Ophthalmoscope, Model: 12500-D, Make: Welch Allyn Inc USA.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Authorization letter is not Embassy Attested.

**Technical Evaluation Criteria for the Purchase of
Category E Equipment (Section-IV of Statement of Requirement) for the financial year 2017-18**

S#	1	2	3									4								22	23				
			Firm Evaluation Parameters									Product Evaluation Parameter													
			Personnel/ Human Resource			Workshop facility Testing/Calibration of Equipment			Manufacturer Performance			Conformanc e To Specification		Requisite Certifications				Product Performance Index							
			Audited Balance Sheets & Income Tax return for last three years (each year 2 marks with Sales above Rs.200.0 million per year)	Financial Soundness exhibiting last three years bank statement showing capability to execute Contract. (Rs.100 M/PKR or above) (03 marks per certificate)	Minimum 03 contracts of fabrication completed. (Min. Rs.200.00 M or above (02 points per contract. Copy of work order be provided)	35 Facility at Peshawar- Khyber Pakhtunkhwa	After Sales satisfactory certificate from at least 03 users of same items (Per certificate 2 Nos.)	Detail of Spare parts availability on firm's pad (for the quoted items)	Mobile After Sales Vans in Peshawar for onsite repair & maintenance of Vehicles & Equipment.	Valid ISO 14001 (Environment Certificate)	Valid ISO 9001 Quality Management Certificate	Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four Minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted)	Authorization Certificate of Manufacturer if the bidder is not a Manufacturer (i.e. Chassis / Superstructure). The manufacturer should meet all the criteria given in these documents.)	Registration with PEC in relevant category i.e. fabrication & Code i.e. MEO6 (Up to Project Value of 300 Million)	Valid Registration with Sales & Income Tax Authorities	Weightage for local Pakistani original manufacturer (OEM certificate from PEC or any other relevant body)	Minimum 05 years of successful fabrication. (Satisfactory performance certificate from at least 03 users. (2 marks per certificate)	Warranty Period three years both with spare parts and services & Next two year services without parts, from the date of handing/taking over	Total			Status	Remarks		
—	—	—	6	6	9	6	6	5	8	3	3	30	No marks, being mandatory parameter				6	6	6	100					
1	Toyota Khyber Motors	Ambulances	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Not registered with PEC in ME06 code. Catalog ambulatory items not attached. Also make/model of ambulatory items not available.
	Colibrative Heavy Industries (Pvt) Limited	Ambulance Hino Dutro WU-640R, 4 x 2, Origin Japan Assembled in Pakistan.	6	6	9	6	6	5	8	3	3	30	Y	Y	Y	-	6	0	88	Responsive	Nil				
	Khawaja Brothers	TOYOTA Hiace Dual A/C High Roof 2.5L	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not from Indus Motors (2) fabricator not register with PEC in the relevant code ME06	
	Colibrative Heavy Industries (Pvt) Limited	Ambulance Toyota Hiace High Roof (4x2) with fabrication and ambulatory items	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not from Indus Motors
2	Colibrative Heavy Industries (Pvt) Limited	Ambulance Toyota Hilux 4x4 single cabine with fabrication and	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not from Indus Motors
	Khawaja Brothers	TOYOTA Hilux 4x4 Single Cabin Pickup	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not from Indus Motors (2) fabricator not register with PEC in the relevant code ME06

**Technical Evaluation Criteria for the Purchase of
Category F Equipment (Left over items) (Section-IV of Statement of Requirement) for the financial year 2017-18**

#	No of Firms submitted Bids	Firms submitted Bids	Name of equipment	Firm Evaluation Parameters									Product Evaluation Parameter						Status	Remarks
				Personnel/ Human Resource		Workshop facility Testing/Calib ration of Equipment		Manufac turer Performance		Conformance To Specification			Product Certification			Product Local Performance				
				Audited Balance Sheets - Income Tax return & Sales tax returns for last three years 03 marks for each year (Rs 1.00 Million or above)	Mobile after sales vans in Peshawar for on sight repair	Local Training of Technical Staff (02 person)	Availability of workshop in Khyber Pakhtunkhwa	Availability of workshop at National level	Valid ISO 14001 (Environment Certificate)	Valid ISO 9001 - 2008 Quality Management Certificate	Fully compliance with the required specifications as per statement of Requirement (Up to a maximum of four minor deviations may be accommodated subject to the condition that main function and performance in any aspect would not affect. However, up to four marks will be deducted)	Full load test report from manufacturer	Authorization certificate from the manufacturer if the bidders is not a manufacturer	Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy)	In case of manufacturer provision of OEM certificate. OEM means that whole unit must be packaged by the manufacturer of the engine & no third party packaging is allowed.	Registration with PEC in relevant category & Codes (Compressor generator & MED3), Applicable in case of authorized agent	Down time for response to the fault	02 working days from first communication from PE		
				9	7	4	8	6	3	3	40	6	No marks, being mandatory			10	4	100		
1	IBS Pharmaceutical	40 KVA Generator. Model: XHX-C40 Make: Fujian Xinheng Xin Motors China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided.
2	Khawaja Brothers	Generator 40KVA, with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-40(PERKINS 1103A-33TG1) Make Erga Genset Spain	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not embassy attested
3	MIA Energy (PVT) LTD.	40 KVA Generator. Model: KGP45.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Authorization not Embassy Attested. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided
4	Millat Tractors Limited	40KVA Generator	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents i.e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. (2) No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
5	Orient Energy	40KVA Diesel Generator. Model: C44D5. Make:Cummins, UAE	9	0	4	8	6	2	2	40	0	Yes	Yes	-	Yes	10	4	85	Responsive	Nil

6	POWER ZONE	GENERATOR 45 kVA MODEL: -PZ 50 MAKE : NOT MENTIONED	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	1.Authorization for engine is from third party that is Mayphil in stead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
7	Sherazi Trading Co Pvt Ltd	40KVA Engine Model: Cummins 14BT3.9G2 Alternator: Stamford Control Module: DSE-6020 Powered by AKSA Turkey From China Plant	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
1	IBS Pharmaceutical	60 KVA Generator. Model: XHX-C60 Make: Fujian Xinheng Xin Motors China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided.
2	Khawaja Brothers	Generator 60KVA, with Parkins UK Engine and MECC- ALTE Alternator Genset Model ERP-60S(PERKINS 1103A-33TG3) Make Erga Genset Spain	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not embassy attested
3	MIA Energy (PVT) LTD.	60 KVA Generator. Model KGP60.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Authorization not Embassy Attested. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided
4	Millat Tractors Limited	60KVA Generator Model: 1103A-33TG2 (Perkins) Alternator: Leroy Somer LSA 42.3	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents i e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested authorization is not available. 3. ME03 not available. 4. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
5	Orient Energy	60KVA Diesel Generator Model:C66DS. Make:Cummins, UAE	9	0	4	8	6	2	2	40	0	Yes	Yes	-	Yes	10	4	85	Responsive	Null	
6	POWER ZONE	GENERATOR 60 kVA MODEL: -PZ 65 MAKE : NOT MENTIONED	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	1.Authorization for engine is from third party that is Mayphil in stead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
7	Sherazi Trading Co Pvt Ltd	60KVA Engine Model: Cummins 4BT3.9G2 Alternator: Stamford Control Module: Deepsea module Powered by AKSA Turkey From China Plant	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.

3	1	MIA Energy (PVT) LTD.	100 KVA Generator. Model KGP100.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Authorization not Embassy Attested. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided
	2	IBS Pharmaceutical	100 KVA Generator. Model: XHX-C100 Make: Fujian Xinheng Xin Motors China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided.
	3	Khawaja Brothers	Generator 100KVA with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-100(PERKINS 1104C-44TAG2) Make Erga Genset Spain	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not embassy attested
	4	Millat Tractors Limited	100KVA Generator Model: 1104A-44TG2 (Perkins) Alternator: Leroy Somer TAL040-TAL042-TAL044	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents i.e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested
	5	Orient Energy	100KVA Diesel Generator Model: C110D5. Make: Cummins, China.	9	0	4	8	6	2	0	40	6	Yes	Yes	-	Yes	10	4	89	Responsive	Null
	6	POWER ZONE	GENERATOR 100 kVA MODEL: -PZ 100 MAKE : NOT MENTIONED	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	1.Authorization for engine is from third party that is Mayphil instead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
	7	Sherazi Trading Co Pvt Ltd	100KVA Engine Model: Cummins 6BT5.9G2 Alternator: Stamford Model: UCI274C Control Module: Deepsea module Powered by AKSA Turkey From China Plant	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
	1	IBS Pharmaceutical	200 KVA Generator. Model: XHX-C200 Make: Fujian Xinheng Xin Motors China	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Registration with PEC in ME 03 code is not provided. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided.	
	2	Khawaja Brothers	Generator 200KVA with Parkins UK Engine and MECC-ALTE Alternator Genset Model ERP-200(PERKINS 1106A-70TAG4) Make Erga Genset Spain	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non responsive	Authorization not embassy attested	
	3	MIA Energy (PVT) LTD.	200 KVA Generator. Model KGP200.	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	(1) Authorization not Embassy Attested. (2) Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) not provided	

4	Millat Tractors Limited	200KVA Generator Model: 1106A-70TG4 (Perkins) Alternator: Leroy Somer TAL046-TAL047-TAL049	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents i.e (1). Technical offer containing details i.e construction of foundation pad, ATS Panel, weather and sound proof canopy, installation and testing of the unit, earthing network, gauge and description of cable not available. 2. Embassy Attested authorization is not available. 3. ME03 not available. 4. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
5	Orient Energy	200KVA Diesel Generator Model: C220D5. Make:Cummins, China	9	0	4	8	6	2	0	40	6	Yes	Yes	-	Yes	10	4	89	Responsive	Nil			
6	POWER ZONE	GENERATOR 200 kVA MODEL: -PZ 200 MAKE : NOT MENTIONED	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	1.Authorization for engine is from third party that is Mayphil instead of Perkins. 2.Authorization for alternator is not available. 3. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.
7	Sherazi Trading Co Pvt Ltd	200KVA Engine Model: Cummins 8CTAA8.9G2 Alternator: Stamford Control Module: Deepsea module Powered by AKSA Turkey From China Plant	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	Non Responsive	Mandatory documents, 1. Embassy Attested authorization is not available. 2. No Service Manuals with circuit diagram for repair and maintenance of Electrical/Mechanical (Soft & Hard Copy) available.